

Jet Ventilator Sterilization Guidelines (Instrumentation Industries)

Sterilization Guidelines for Jet Ventilator:

- **The exterior of the on/off valve and high pressure hose should be wiped off with a disinfectant.**
- **Do not autoclave or immerse in fluids**
- **The small bore tubing assembly is for single use only**