

Ventilator Tubing Sterilization Guidelines (Instrumentation Industries)

Ventilator Tubing

Tubing should be cleaned with a mild detergent followed by a sterile water rinse prior to processing.

- Steam Autoclave at 270°F (132°C)
- Pasteurize at 170°F (77°C)
- EtO at 131°F (55°C)
- Gamma Radiation